

BurButler™

Unique Silicone Block
Grips All Burs

Instructions for Use

BurButler silicone bur blocks accommodate CA, FG, and short shank rotary instruments. Mix and match the shanks to suit your clinical needs. Assorted colors allow you to organize and care for your dental burs making your practice more efficient. BurButler bur blocks are reusable and fully autoclavable.

General Instructions

Follow all manufacturers' instructions for proper disinfecting and autoclaving procedures for all rotary instruments.

First process through an ultrasonic cleaner or a washer disinfectant. Once dried place the assembled BurButler into a sterilization pouch or wrap.

Ultrasonic Cleaning

Prepare a cleaning solution following the manufacturers instructions.

Place the BurButler with the lid removed in a sonic unit ensuring the device is completely submerged for at least 15 minutes.

Finish by rinsing under running water thoroughly for 1 minute.

Visually inspect to confirm removal of any debris.

CAUTION: Always follow manufacturer's protocols for effective cleaning. Do not rest the BurButler on the tank bottom. Do not overload, or stack heavy or bulky items on the surface of the BurButler's plastic lid as doing so may damage the lid.

Washer Disinfectant

We recommend that the lids are washed separate from the block at this stage. Do not invert the lids. This allows adequate drainage.

Once the cycle has been completed the lids can be placed onto the BurButler bases.

Continued on the other side.

Autoclave

The base and lid should be assembled and placed in sterilization pouches prior to autoclaving.

Follow the manufacturer's instruction as recommended to achieve adequate sterilization. Do not exceed 134°C (270°F) as doing so will damage the lid.

CAUTION: Always follow manufacturer's protocols for effective sterilization. Do not wrap the BurButler too tightly. Do not overload, or stack heavy or bulky items on the surface of the BurButler's plastic lid as doing so may damage the lid.

Note On Use

The force that you insert the bur will be the force required to remove the bur. Do not jam the bur in too tightly as you will have to pull equally to remove it.

BurButler must be thoroughly cleaned prior to the first use and prior to each subsequent reuse.

Do not use if drainage holes are obstructed.

Wear appropriate personal protective equipment when using the BurButler.

BurButler silicone is stain resistant. If a metal bur corrodes and marks the surface, this will wash off with a soft cloth and ultrasonic cleaning. Do not insert damaged burs and this will not occur.

For Bacteriological report please see our website www.shofu.com

BurButler

Unique Silicone Block
Grips All Burs

Distributed by

Shofu Dental Corporation

1225 Stone Drive, San Marcos, CA, USA

www.shofu.com • (800) 827-4638

8210DFU v2
1115